

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Serum - plasma- milk	Detection of Brucella using: BAPA, Rose- Bengal, milk ring, Rivanol precipitation, Serum agglutination, Indirect ELISA, competitive ELISA and Complement fixation	OIE Terrestrial Manual : 2016 Chapter 2.1.4	Balance HF-300GDPM600 S.N.: 1283478S.N.: G89790 Benchtop centrifuge Harmonic Series7000599 S.N.: 812037S.N.: 180750 Magnetic stirrer, S-180520 S.N.: 2711612 Incubator, P selecta S.N.: 186789 Autoclave, KT/40 S.N.: 801552 Card rocker, Cr1 S.N.: 25 Vortex mixer, S200-230V-EU S.N.:Z807110 Centrifuge igma 3-16P80-2 S.N.:131449Code: 9

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Parenteral sterile pharmaceutical and biological products	Rabbit pyrogen test	International , British, USP and European pharmacopoeia 2017	Eval Flex: S.N. :13835 Weighing scale, JZC-TSED-105 S.N.: 80715233 Incucell Incubator, Incuce 11-111 S.N.: B070756
Pharmaceutical products	Assay of tetracyclines by HPLC	In house method 2.(1.2.1)	HPLC Agilent technologies 1200 Series S/N:JP94177309 DE: 62976281 DE: 64779187 DE: 90374827 DE: AAX00739 Balance: Chyo Js-110 119966

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Food and animal feeding stuffs	Enumeration of total aerobic plate count	American Public Health Association - APHA:2001 ISO 4833: 1-2:2013	Biosafety cabinet class II Model: Nu.425-400E Model: 810-11-AS.N.: 83835 S.N.: 20669
	Enumeration of <i>Enterobacteriaceae</i>	National Standard Method - APHA:2001	Autoclave Model: Raypa AES Model: HL-340 S.N.: 36709 S.N.: 0002013
	Count and isolation of <i>Staphylococcus aureus</i>	-FDA:BAM: chapter 12; 2001	Incubator Model: MmoistureMM IncucellS.N.: B030203, B040506
	Enumeration of <i>Escherichia coli</i> and the coliform bacteria	-FDA: BAM: chapter 4; 2002	Model: Haraveus B12 S.N.: 94109525, S.N.: 98107789 S.N.: 98107326, S.N.: 98105793
	Detection of <i>Listeria monocytogenes</i>	FDA: BAM: chapter 10;(2016)	Oven Model: Fisher Model: Venticell S.N.: 250819 S.N.: B030778, B030779
	Preparation ,production and assuring of microbiological culture media	ISO 11133-2014	pH meter, Model: 3510 S.N.: 04528 Balance Model: AJ-620E Model: GB4102 S.N.: 083530157 S.N.: 12808049

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Food and animal feeding stuffs	Detection of <i>Vibrio parahaemolyticus</i>	ISO 21872-1:2017	Balance Model Mettler, S.N G89790 Autoclave S.N 4984 Incubator Model IB-05/15625G S.N R049042 Oven Model ON-02G/12G22G S.N R059016
	Enumeration of yeasts and moulds	ISO 21527 -1&2: 2008	
	Detection of <i>salmonella</i> spp.	ISO 6579: 2017	
	Enumeration of β glucuronidase-positive E.coli colony count technique at 44°C using 5-bromo-4chloro-3indolyl β Dglucouronide	ISO 16649-2:2001	
	Detection of <i>Shigella</i> spp.	ISO 21567:2004	
Water	Enumeration of culturable micro-organisms- Colony count by inoculation in a nutrient agar culture medium	ISO 6222/ 1999	
	enumeration of Escherichia coli and coliform bacteria	FDA: 2002	

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Meat	Differentiation between animal meat species using PCR	Identification of Meat Species by Polymerase Chain Reaction (PCR) Technique:2007	Real time PCR Machine T1Gene Amp 9600 S.N.: 1203259 S.N.: P3831
	Detection of foot and mouth disease virus (FMDV) by real time PCR (qRT-PCR)	In-house method based on OIE, 2017	Hybridization incubator, Lab –line 308-1 S.N.: 08930288
	Detection of bovine viral diarrhea virus genotype 1 by real time PCR (qRT-PCR)	-- In-house method based on OIE, 2015	Balance Model: A-160, S.N.: 0053626
	Detection of mycobacterium avian subsp. Para tuberculosis (MAP) by real time PCR (qRT-PCR)	- In-house method based on OIE, 2014	Biosafety cabine classll, Model: Auramini, S.N.: Jo9b0N3064, Jo9b0N3060 Biosafety cabine classl, Model: Nu-425-4005, S.N.: 56818 Microwave, RS561W S.N.: 9403156623

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Serum	Detection of antibodies of Theileria (Babesia) equi in equine using cELISA	OIE Terrestrial Manual :2014 Chapter 2.5.8	Centrifuge Model: 203 S.N.:39952
Serum	Detection of antibodies of Babesia caballi in equines using cELISA	OIE Terrestrial Manual :2014 Chapter 2.5.8	Digital thermohygrometer Model: HI 9565 S.N.: 08382928
	Detection of Dourine	OIE – Chapter 2.5.3.(2013)	Microplate reader Model: Multiskan FC S.N.: 357-902765 Centrifuge Model: 203 S.N.:39952

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Milk and milk product	isolation and identification of pathogens causing mastitis.	- National standard method : staining procedures (NSM 2007).	Vitek2 compact Model: 15 S.N:12170
		- US enviromental projection agency office of pesticidal programs 2014	Incubator Model: MH60 S.N:41564298 Refrigrator 1-Model: N/A S.N:NE33SN 2- Model: N/A S.N:K33M Microscope S.N:3151000130 Safety cabinet S.N:603851 Centrifuge Model: Z326K S.N:66130271

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Serum	Capture ELISA for bovine viral diarrhoea BVD antigen detection (BVDV antigen ELISA test kit)	OIE Terrestrial Manual Chapter 2.4.7:2015	ELISA reader Model: Elx808 S.N: 218339
	Detection of non-structure protein of FMDV using (ELISA)	OIE Terrestrial Manual :2017 Chapter 2.1.8	Incubator Model: venticell S.N: 060655
	Blocking ELISA for detection of antibodies for bovine leucosis virus (BLV)	OIE Terrestrial Manual :2012 Chapter 2.4.10	Model: Shell lab2406-2, SN.: 12018111
	Equine Infectious Anemia antibodies Detection in serum (EIA) by agar gel immune diffusion AGID	OIE Terrestrial Manual :2013 Chapter 2.5.6	Balance Model: Chyo mj-500 S.N: 121534
	Detection of antibodies against Blue Tongue disease virus by ELISA	OIE – chapter 2.1.3 (2014)	Oven Model: haereus S.N: 12880
	Detection of antibodies against Enzootic Bovine Leukosis by AGID	OIE- chapter 2.4.10 (2012)	Autoclave Model Jsac-04 S.N: 080506
	African Horse Thickness AHS antibodies detecting ELISA Test	OIE- chapter 2.5.1 (2017)	
	Equine Viral arteritis EVA antibodies detecting ELISA Test	OIE- chapter 2.5.10 (2013)	
	Equine Herpes virus EHV antibodies detecting ELISA Test	OIE- chapter 2.5.9 (2017)	
	Detection of Equine Influenza by ELISA test	OIE- chapter 2.5.7(2016)	
	Detection of West Nile Fever by ELISA test	OIE- chapter 2.1.24 (2013)	
	Detection of vesicular stomatitis	OIE- chapter 2.1.23 (2015)	
	Detection of RVF-antibodies IgM, IgG ELISA test	OIE- chapter 2.1.18 (2016)	

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Cattle serum and milk	Detection of <i>Mycobacterium Paratuberculosis</i> antibodies by Indirect ELISA	OIE manual :2014 Chapter 2.1.11	Microplate Reader Model:MR-96 SN MR2DO003
Tissue samples and isolated bacterial culture	Detection of mycobacterium tuberculosis complex by RT-PCR	In house method based on the in use kits and OIE- chapter 2.4.6 (2009)	Incubator Model:Velp FOC225E SN F10300170 Real Time PCR SN. 271003107 Model: Step one Laminar air flow SN.: PCB1200-1207103 Model: Cryste Puricube Incubator SN.: BT1020 Model: BTC Centrifuge SN.: 141998 Model: Sigma1-14

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Swab samples or internal organs	Equine microbiology	UK standers 2014 – 2015 with different IDs OIE manual section 2.5; 2012 OIE manual 2.9.8; 2016	Incubator Model: IB-OSG S.N:R049041 Balance Model: Preciso1212M S.N:59127
	Glander	OIE- chapter 2.5.11 (2012)	
Food and Animal feeds	Enumeration of <i>Clostridium perfringens</i> -colony count technique	ISO 7937 : 2004	Incubator Model IB-05/15625G S.N R049042 Oven Model ON-02G/12G22G S.N R059016

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Food and feedstuffs	Horizontal method for detection of <i>Salmonella</i> spp.	ISO 6579: 2002/Amd:2007	Balance Model YMCCO, S.N 121520 Incubator Model En 500 S.N 01-157 Model BDS (20-256) S.N 1206B05DD
<i>Salmonella</i> strains on TSI or nutrient agar	Serotyping of <i>Salmonella</i>	ISO/TR 6579- part 3: 2014	Thermo- freezer S.N 811775-50004 Oven Model Heratherm OMS180 S.N 41472199 Incubator Model IGS60 S.N 41377023

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Serum	Serological Diagnosis of Contagious Bovine Pleuropneumonia (CBPP) (Antibody detection of Mycoplasma)	OIE Manual: 2014 Chapter 2.4.8 OIE Manual: 2017 Chapter 2.3.5	Microplate ELISA Reader S.N UV 28266
Tracheal & nasal swabs, mik, lung tissue, liver, spleen, lymph nodes, eggs, chicken, semen, placenta, fetal 7 vaginal discharge, tissue cultures and vaccines	Molecular diagnosis of Mycoplasma DNA using PCR	OIE Manual: 2017 Chapter 2.3.5	PCR Thermo-cycler Model: S1000 Biorad S/N: SC007858

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Serum	Detection of antibodies against peste des petiltes Ruminants(PPR) Virus	OIE Manual: 2013 Chapter 2.7.10	ELISA Reader Model: SN:357-902756 Incubator Model: SN:
Fish	Bacterial diseases in fish	National Wild Fish health survey laboratory procedures manual, 2009	Oven Model: Heraeus T-50, SN.: 5003154 Incubator SN.: B042306 Model: MedcenterEimriclungen GumbH Balance SN.: 2200105029164 Model: ScaltecSBA3IMAX220C Lamina Flow SN.: 15255 Model: Tel-star Bio-11-A
	Detection of potentially enteropathogenic <i>Vibrio</i> spp.	ISO/TS 21872-1 (2007)	
Meat and Poultry Tissue	Detection, Identification and Quantitation of Antimicrobial Residues	USDA- CLG-2011	

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Water	Determination of pH	AOAC: 2016	pH meter Jenway 3510 Magnetic stirrer SN.: 9249 Model: Pyromulti Lab-Line 1963-1
	Determination of sulfate	AOAC: 2016	Spectrophotometer Colorometer SN.: 1240 Model: Jenway 6030 Spectrophotometer S.N.:100126 Model:Cocil England
	Assay of heavy metals by Atomic absorption	AOAC: 2016	Atomic Absorption Model: SenseAA S.N: A.C.N.005 472 686 Balance Model No: SA310, S.N.: 9901
	Metal	LOQ (ppm)	
	Cu	0.05	
	Mn	0.04	
Na	0.10		
Zn	0.14		

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
milk and milk products	Determination of aflatoxin M ₁	ISO 14675:2003	ELISA Reader Model: infinite F50 S.N.: 1002002254 RidA screen Aflatoxin M1 Art No. R 1121 Microtiter plate spectrophotometer SN.: 1002002254 Model: Infinite F50 Balance Model: MXX-612, S.N.: 17852672 Model: VIC 1501, S.N.: 24503958 Incubator Model: Heraeus, S.N.: 50042307
Meat and Meat product	Determination of trenbolone by ELISA	USDA- CLG -2004	
	quantitative analysis of ractopamine by ELISA	RIDASCREEN Ractopamin (Art. No.: R9901)	
Fish	Screening for Malachite Green and Crystal Violet by ELISA	CLG-MGCV1.02-USDA,2016	

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Food and Feed	Determination of fat %	ISO 1443:1973	Oven Model: Venticell S.N.:007151 Balance Model: SCINETECH S.N.: 9895 Automatic extractor ModelSER148 S.N.: 279716
		Official Journal of the European Communities commission directive 98/64/EC (1998)	
		FAO 14/7: (1986).	
Food and Feed	Determination of ash %	AOAC 938.08:2016 (Seafood)	Balance Model SCINETECH S.N.: 9895 Muffle furnace S.N.: 35
		AOAC 920.153 :2016 (Meat)	
		AOAC 945.46 :2016 (Milk)	
		AOAC 920.108 :2016 (Cream)	
		AOAC 920.115 :2016 (Sweetened condensed milk)	
		AOAC 930.30 :2016 (Dried milk)	
		AOAC 935.42 :2016 (Cheese)	
		AOAC 942.05 :2016 (Animal Feed)	
AOAC 945.48 :2016 (Evaporated milk unsweetened)			

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Food and Feeds	Determination of moisture % and total Solids%	AOAC 930.15 :2016 (Feed)	Hot Air Oven Model: Venticell S.N.:007151 Balance Model: SCINETECH S.N.: 9895 Oven: model: V0S-301SD S.N.: 11212728
		AOAC 950.46 :2016 (Meat)	
		AOAC 920.116 :2016 (Butter)	
		AOAC 926.08 : 2016 (Cheese)	
		AOAC 927.05 : 2016 (Dried milk)	
		AOAC 941.08 :2016 (Ice cream)	
		AOAC 952.08 :2016 (Seafood)	
		AOAC 945.48: 2016 (Evaporated milk (unsweeted))	
		AOAC 920.115: 2016(Sweetened condensed milk)	
AOAC 990.20 :2016 (Total solid in Milk)			
Fish and fish products Meat and meat products	Determination of Total Volatile Basic Nitrogen	Egyptian Standard 2760-1/2006 Egyptian Standard 63-9/2006 Commission Regulation (EC) No 1022/2008 of 17 october 2008	Balance Model: SCINETECH, S.N.: 9895 Model GP 4102, S.N.:12808038 Kjeldahl Distillation Unit Model:VDK-129-VELP, S.N. 358985 Mentel heater

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

Materials/Products Tested	Types of Test/Properties Measured /Range of Measurement	Standard Specifications / Techniques Used	
Food and Feeds	Determination of crude protein	AOAC 981.10 :2016(Meat)	Balance Model: SCINETECH SN.: 9895 Kjeldahl Digestion Unit Model:DK6-VELP S.N.:251476 Model:DK-129-VELP S.N.:30769
		AOAC 991.20 :2016(Milk)	
		AOAC 920.109:2016(Cream)	
		AOAC :945.48:2016(Evaporated milk (unsweetened))	
		AOAC :920.115:2016(Sweetened condensed milk)	
		AOAC :930.29:2016(Dried milk)	
		AOAC :2001.14:2016 (Total nitrogen in Chesse)	
		AOAC 920.123:2016 (Cheese)	
Milk and milk products Honey	Determination of titretable acidity	AOAC 947.05, 920.124, 962.19: 2016	Balance Model: SCINETECH, S.N.: 9895 pH meter model 3310, S.N. 5690
		ISO 6091/2010 and ISO 6092/1980 FAO, 1997	
Milk and milk products	Determination of Lactose%	AOAC 923.09 and 930.28(2016) milk National food safety standard, GB 5413.5-2010	Balance Model: SCINETECH, SN.: 9895

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Animal Health Research Institute Laboratories
Animal Health Research Institute (AHRI) - Main Branch
Nadi Elsaid Street, Dokki
Giza – Egypt

Schedule No.: 206106B Issue No. (5): November 8, 2018 1st Accreditation Date: October 8, 2010 Revision No. (): Valid to: October 31, 2020

<i>Measured Quantity</i>	<i>Range</i>	<i>Calibration and Measurement Capability*</i>	Brief Description of Measurement and Equipment Used	
Calibration of temperature indicators	From - 40 °C to 50 °C From 50 °C to 180 °C	± 0.16 °C ± 0.6 °C	In house method AHRI - 1.13.1 Cal based on ES 651:2002 ES 3742:2002	Thermocuple Model: CTH-609 PT 100 Thermocuple Type T Digital indicator gluke Model: 1529 S.N: B36900