

Schedule of Accreditation

Food and Water Labs for Chemistry , Microbiology , Pesticide residue , Mycotoxins and heavy metal Analysis
Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street – Abdeen - Cairo – Egypt

CAB No.: 208002 B Issue No. (4): July 10, 2017 1st Accreditation date: January 12, 2009 Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement	Standard Specifications/ Techniques Used
Food	Detection of <i>Salmonella</i>	ISO 6579:2017
	Detection of <i>Salmonella</i> by using VIDAS	VIDAS REF 30 702- 2013 06984T -en-201307
	Detection of <i>Listeria monocytogenes</i>	ISO 11290-1:2017 VIDAS REF 30 704 LMO2: 2010
	Enumeration of total bacterial count	ISO 4833-1 : 2013
	Enumeration of coagulase positive <i>Staphylococcus</i>	ISO 6888-1 :2003
	Enumeration of total coliform group count	ISO 4832 : 2006
	Enumeration of β -glucuronidase-positive <i>E. coli</i> count	ISO 16649-2 :2001
	Detection of <i>Enterobacter sakazakii</i>	ISO 22964 :2006
	Detection of <i>E. coli</i> O157 : H7	VIDAS REF 30122 ECP1 2011
	Enumeration of <i>Bacillus cereus</i>	ISO 7932:2004
		OVEN S.N.L099003 OVEN S.N:01-0059 Incubator Model Memmen, S.N 780505, 780523, 780515, 280503, e599-0392 Model Lab. Tech, S.N.: 07118314, 071118312 Heraeus, S.N 96107345, . 96107903, 961093361, N97112508,96107897, 406697400 Model LTE, S.N 47631/6 Model Jlabtech S.N071118311 Balance Model OHAUS, S.N 1203220051-112220716 Model Seoul, S.N 7129151697 Autoclave Model Selecia SN CHL-0107-T Model Hirayama SN 96108542 Model HV-85, S.N5538 ,3495 VIDAS Model BioMerieux S.N IVD 3004087

Kornish El-Masdi, Rind El-Masdi Tower 1, Cairo – Egypt

Tel.: (002) 25275220/25275224/25275225/25275227

Fax: (002) 25275226

F10P15G

5.2/Os1

2015

Page 1 of 6

Hany Desouky

كامله القوي - مع رند القوي - القوي - القوي
القوي الوطني للاعتماد / القوي الوطني للاعتماد
القوي الوطني للاعتماد

Schedule of Accreditation

Food and Water Labs for Chemistry , Microbiology , Pesticide residue , Mycotoxins and heavy metal Analysis
Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street - Abdeen - Cairo - Egypt

CAB No.: 2089002 B Issue No. (4): July 10, 2017

1st Accreditation date: January 12, 2009

Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured/ Range of Measurement	Standard Specifications/ Techniques Used
Water (Renal dialysis and well water)	Detection and enumeration of coliform Group	SM 9222B :2012 Incubator Model Jabtech . 071118311
	Heterotrophic plate count in water at 35°C	SM 9215B :2012 ModelMettmert, c599-0392
	Detection and enumeration of enterococci	SFS-ISO 7899-2: 2000 ModelHerarus 96107897, 40697400, 97112508
	Detection and Enumeration of <i>Pseudomonas aeruginosa</i> method by membrane filtration.	ISO 16266:2006 OVEN Model Binder SN 86241, Model Imperial SN 60160.1491.
	Detection and Enumeration of <i>E.coli</i>	SM 9222D :2012 Autoclave Model Sebecia SN CHL-0107-T Model Hirayama SN 96108542 Model HV-85, S.N5538 , 3495


Hany El Desouki

Schedule of Accreditation

Food and Water Labs for Chemistry , Microbiology , Pesticide residue , Mycotoxins and heavy metal Analysis
Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street - Abdeen - Cairo - Egypt

CAB No.: 208002 B Issue No. (4): July 10, 2017

1st Accreditation date: January 12, 2009

Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement	Standard Specifications/ Techniques Used	AliaTest Fluorometer Manufacturer: VICAM, USA Model series IV, S.N 0554 Balance Manufacturer: Sartorius, Canada Model BP410, S.N 60707752 Muffle Furnace Model ELF 11/14B, S.N 21-200701 Model 48000, S.N 1166000105409 Model ELE 11/14B, S.N 21-201992 Model M-52S, S.N CKN9440-124 Oven Model OV-11, S.N K-119001 Model N.A, S.N 930001 Digester Model F50100200, S.N 230004 Model F50100200, S.N 203774 Model 10014134, S.N 52009128 Spectrophotometer Model UV LAMB, S.N 501S12052005 Balance Model BP410, S.N 60707752 Balance Model COBOS, S.N, 083900249
Pearl, Corn Pistachio, Chilli Powder & Dried figs	Determination of total aflatoxin (B1, B2, G1, G2)	AOAC 991.31 : 2016	AliaTest Fluorometer Instruction Manual
		AOAC 923.03 :2016	
Corn Flour /Wheat Flour	Determination of ash content	AOAC 941.12: 2016	
Spices	Determination of moisture	ES 465-3: 2005	
Cocoa and its products	Determination of total nitrogen& total protein	AOAC 981.10:2016	
Meat	Determination of total nitrogen	AOAC 991.20:2016	
Milk	Determination of Hydroxy methyl furfural	AOAC 980.23:2016	
Honey	Determination of Peroxide value	ES 355-2:2005	
Fats & Oil	Determination of acid value and acidity	AOAC 965.33:2016 In-house method F.Ch-11 based on ES-51- 2:2005	
		ISO 660/2009	

Schedule of Accreditation

Food and Water Labs for Chemistry , Microbiology , Pesticide residue , Mycotoxins and heavy metal Analysis
Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street - Abdein - Cairo - Egypt

CAB No.: 208002 B Issue No. (4): July 10, 2017

1st Accreditation date: January 12, 2009

Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement	Standard Specifications/ Techniques Used
Soft drinks	Determination of Benzoic acid	In-house method F.Ch-13 based on J. Medicinal Plants:2010
Food Sudan dyes, I, IV quantitative and II,III (qualitative)	Determination of Color	ASTA 28.0- 2013
Acidified food	Determination of pH	AOAC 981.12-2016
Fish	Determination of Total volatile basic nitrogen	ES 2760-1- 2006
Milk and milk product	Determination of Total fat	ES 155:1974

Highly El Desouki



Schedule of Accreditation

Food and Water Labs for Chemistry, Microbiology, Pesticide residue, Mycotoxins and heavy metal Analysis

Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street - Abdeen - Cairo - Egypt

CAB No.: 208002 B Issue No. (4): July 10, 2017

1st Accreditation date: January 12, 2009

Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement	Standard Specifications/ Techniques Used
Food (Fruits, Vegetables, Cereal)	Determination of Pesticide	AOAC 2007.01:2016 QUECHERS Method
	Chlorpyrifos Me	GC/MS Manufacture: Agilent Model: GC:7890A-MS:5973C S.N :GC:US:1224026
	Primapthos Me	
	Malathion	
	Dieldrin	
	P,P DDE	
	O,P DDT	
	Endosulfan sulfate	
	Trifloxystrobin	
	Piperonyl butoxide	
	Bifenthrin	
Plant origin food	Determination of Trace elements	AOAC 999.10-2016NMKL
	Metal	
	Pb	
	Cl	
	Cu	
	Zn	

Hanna El Desouki
كوبرك الحادي - جى يحيى الحادي - القلي - مصر
قبة (19999999) / قبة (19999999) / قبة (19999999) / قبة (19999999)
قبة (19999999)

Schedule of Accreditation

Food and Water Labs for Chemistry , Microbiology , Pesticide residue , Mycotoxins and heavy metal Analysis
Central public Health Laboratories- Ministry of Health and Population
19 El Sheikh Rehan Street - Abdein - Cairo - Egypt

CAB No.: 208002 B Issue No. (4): July 10, 2017

1st Accreditation date: January 12, 2009 Revision No():

Valid to : January 11, 2021

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement	Standard Specifications/ Techniques Used		
Drinking Water	Determination of metals	SM 3120 B:2012 EPA 200.7: 1982		
			Metal	LOQ (ppm)
			Cu	0.0011
			Fe	0.0039
			Mn	0.00049
			Zn	0.0028
			Na	0.801
			K	0.301
			Mg	0.0149
			Ca	0.195
Cr	0.0035			
Ni	0.0034	ICP Spectrometer Model Optima 7300DV S.N 077C0052501.		

Mary El Desouki