

**Schedule of Accreditation**  
for Testing Laboratory According to ISO/IEC 17025  
issued to

**Ethylene, Utility & Environmental Laboratories & Research Sector**  
**Sidi Kerir Petrochemical Company (SIDPEC)**  
36 Km Alexandria - Cairo Desert Road  
El Nahda , El -Meriyah  
Alexandria - Egypt

Schedule No.: 208020 B Issue No. (3): March 15, 2017

1<sup>st</sup> Accreditation date: March 1, 2009

Revision No. (1): March 15, 2017

Valid to: October 31, 2020

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used / Main Device
Water	Determination of pH value	pH Meter Manufacturer: WTW, Germany Model Inolab level 3 terminal S/N: 03440002
	Determination of Electrical Conductivity	Conductivity Meter Model: cond 340i S/N: 04030007
	Determination of Metals	Inductively Coupled Plasma (ICP) Plasma Emission Spectrometer Model: shimadzu ICPS- 7500 S/N: B 40933700208 KG
	Element LOQ (µg/L)	
	Iron 40	
	Cadmium 5	
	Nickel 38	
	Lead 45	
	Copper 28	

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt

Tel.: (202) 25275220/5/6/7

Fax: (202) 25275224

F10P15G

5.3/Mar 2018

Hany A. Pesouki

كورنش الماادي - برج رايض الماادي - القاهرة - مصر

تليفون : ٢٥٢٧٥٢٢٠/٥/٦/٧ (٢٠٢)

فاكس : ٢٥٢٧٥٢٢٤ (٢٠٢)