

Schedule of Accreditation

Issued to

Food & Water Testing Laboratory
Laboratories of Management and Environmental Researches
Agency of Services - Secretariat of Medina Munawara
Ministry of Municipal & Rural Affairs
Madina Munawara – Kingdom of Saudia Arabia

Schedule No.: 210016 B

Issue No. (3): August 10, 2016

1st accreditation date : August 15, 2011

Revision No. (-):

Valid to: August 14, 2019

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement		Standard Specifications/ Techniques Used
	Parameter	(LOQ) (mg/l)	
Bottled & Unbottled Drinking Water	Determination of nitrate	3.00	UV/Visible Spectrophotometer Manufacturer HACH, USA Model DR 4000U S.N 0412U0005935
	Determination of sulfate	3.00	
	Determination of nitrite	0.01	
	Determination of fluoride	0.30	
	Determination of total iron	0.05	

Hany Basouki

Schedule of Accreditation
Issued to
Food & Water Testing Laboratory
Laboratories of Management and Environmental Researches
Agency of Services - Secretariat of Medina Munawara
Ministry of Municipal & Rural Affairs
Madina Munawara – Kingdom of Saudia Arabia

Schedule No.: 210016 B Issue No. (3): August 10, 2016

1st accreditation date : August 15, 2011

Revision No. (-):

Valid to: August 14, 2019

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement (LOQ)		Standard Specifications/ Techniques Used
	Determination of metals	LOQ (µg/l)	
Bottled & Unbottled Drinking Water	Metal		In-house Method TRP-10-07 based on USEPA 200.8 , Rev. 5.4:1994 ICP-MS Model X Series II S.N. SN02039C
	Pb	0.780	
	Cd	0.151	
	Cu	41.500	
	Ba	13.100	
	As	0.215	
	Cr	1.740	

Happy Desouki



Schedule of Accreditation
Issued to
Food & Water Testing Laboratory
Laboratories of Management and Environmental Researches
Agency of Services - Secretariat of Medina Munawara
Ministry of Muncipal & Rural Affairs
Madina Munawara – Kingdom of Saudia Arabia

Schedule No.: 210016 B Issue No. (3): August 10, 2016 1st accreditation date : August 15, 2011 Revision No. (-): Valid to: August 14, 2019

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement (LOQ)*	Standard Specifications/ Techniques Used
Bottled & Unbottled Drinking Water	Detection and enumeration of coliform	SM 9222B:2012
	Detection and enumeration of thermofecal coliform membrane	SM 9222D:2012
	Detection and enumeration of fecal Streptococcus and Enterooccus	SM 9230C :2012
		Membrane Filter Technique Incubator Model G-Cell S.N. 30194 Model MIR 262 S.N. 21103112 Model MIR 262 S.N. 21103110 Model 2001249 S.N. 0432048 Model 2001249 S.N. 557493 Autoclave Model : MLS3780 S.N. 751372 S.N 110072 Oven Model : MOV-212 S.N.40811682 Balance Model 43289 S.N.PD2-5000

Harry El-Souk

Schedule of Accreditation

Issued to

Food & Water Testing Laboratory

Laboratories of Management and Environmental Researches

Agency of Services - Secretariat of Medina Munawara

Ministry of Municipal & Rural Affairs

Madina Munawara – Kingdom of Saudia Arabia

Schedule No.: 210016 B

Issue No. (3): August 10, 2016

1st accreditation date : August 15, 2011

Revision No. (-):

Valid to: August 14, 2019

Materials/ Products Tested	Types of Tests/ Properties Measured /Range of Measurement (LOQ)*	Standard Specifications/ Techniques Used
Fruits and vegetables	Determination of pesticide residues	Gas chromatography
	Pesticide	Model Mass (ISQ)
	LOQ (mg/kg)	S.N : 620110708 S.N: 110825
	Parathion methyl	BS EN 15662:2008
	Malathion	pH meter
	Fenitrothion	Model HQ40D
	Diazinon	S.N 111100063131
	Ethion	Balance
		Model AB104-S
		S.N 1121322949

