

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued to
Central Laboratory for Soil , Water and Plant Analysis
Faculty of Agriculture - Fayoum University

Schedule No.: 212028B Issue No.(2): May 9, 2017 1st Accreditation Date: December 6, 2012 Revision No. (1): November 12, 2018 Valid to: October 31, 2020

Materials / Products Tested	Type of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Water	Measurement of pH value	ES 1842-2 :2007	pH Meter Manufacturer Mettler Toledo, Switzaerland .N 1231215670
	Determination of electrical conductivity at 25 °C	ISO 7888-1985	Conductivity Meter Manufacturer Mettler Toledo, Switzaerland Model Seven Easy S.N 1231225906/EQ021
	Measurment of nitrate as NO ₃ ⁻	LOQ 0.10 mg/l SM 4500- NO ₃ ⁻ B :2012	UV visible spectrophotometer Manufacturer Hitachi, Japan Model U-2900 S.N 22E38-002