



**Schedule of Accreditation**  
for Testing Laboratory A according to ISO/IEC 17025  
Issued To  
**Faqous Laboratory for Drinking Water Analysis**  
Sharkia Potable Water and Sanitation Company (SHAPWASCO)  
Abo Shalaby Station - Faqous  
Sharkia - Egypt

Schedule No.: 212011/1B Issue No. (1): August 16, 2018 1<sup>st</sup> Accreditation date: August 16, 2018 Revision No. (-): Valid to: October 31, 2020

Material / Products Tested	Type of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used
Drinking Water	Determination of pH value	pH Meter SMWW 4500 H <sup>+</sup> B:2017 Manufacturer : USA Model Thermo Orion2Star S.N: B30555
	Determination of Turbidity	Turbidity Meter SM WW2130 B:2017 Manufacturer : HF Scientific, USA Model Micro1000 S.N 16070C034511
	Determination of Conductivity at 25°C	Conductivity Meter SM WW2510 B:2017 Manufacturer : Jenway 4510 Model 4510 S.N: 30435
<b>Determination of Metals</b>		<b>UV- Vis Spectrophotometer</b>
Metals	LOQ (mg/l)	Manufacturer AQUARIUS , Australia
Iron (Fe)	0.068	Model CE7400S
Aluminum (Al)	0.067	S.N 146-456
Manganese (Mn)	0.07	

*M. J. M.*  
*Faqous*

كورنش العادى - برج رياض العادى 1 - القاهرة - مصر  
تليفون : 202 25275220 / 202 25275224  
فاكس : 202 25275224

**Hany N. Dessouky**

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt  
Tel.: (202) 25275220/5/6/7  
Fax: (202) 25275224  
F10P15G  
5.3/Mar 2018  
Page 1 of 1