



Schedule of Accreditation  
Issued To

Irrigation And Hydratics Laboratory  
Faculty Of Engineering - Cairo university  
El Gamaa Street - Giza

Schedule No.: 214010B

Issue No. (I): June 16, 2016

1<sup>st</sup> Accreditation date: June 16, 2016

Revision No. (I): September 10, 2017

Valid to : June 15, 2021

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements (%)	Standard Specifications / Techniques Used
Rotodynamic Pumps <ul style="list-style-type: none"> <li>Submersible pumps</li> <li>Centrifugal pumps</li> </ul>	-Determination of pump total head (H-Qcurve) -Determination of pump efficiency (efficiency curve)	ISO 9906:2012  Electromagnetic flow meter Model : MUT 2200/EL S.N. AAC2588, S.N: AAC2582 Three phase electronic energy meter Model : CE4DMID01 Serial no : 2217860287 Pressure gauge Model: MAN-SD1S Serial no: 408393.1, Serial no : 408393.4

Hany El Desouki

Ras ha

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt

Tel.: (202) 25275220/5/6/7

Fax: (202) 25275224

F10P15G

5.2/Oct 2015

Page 1 of 1

كوريداتى المادى - برج رياض المادى - القاهرة - مصر

تليفون : ٢٥٢٧٥٢٢٠ / ٥ / ٦ / ٧ ( ٢٠٢ )

فاكس : ٢٥٢٧٥٢٢٤ ( ٢٠٢ )