



Schedule of Accreditation
 Issued To
Radiometry Lab
National Institute For Standards (NIS)
 Tersa Street El-Haram
 Giza - Egypt

Schedule No.:215019B

Issue No.(1) : September 5, 2016

1st Accreditation Date: September 5, 2017

Revision No. (1): August 28, 2017

Valid to : September 4, 2020

Materials / Products Tested	Types of Tests / Properties Measured / Range of Measurements (%)	Standard Specifications / Techniques Used
Phototherapy Luminaire	Determination of the total Irradiance level Irradiance level 200-1000 $\mu\text{W}/\text{cm}^2$ (5-23 $\mu\text{W}/\text{cm}^2, \text{nm}$)	In house method 52-TM-T2 based on BS EN 60601-2- 50:2009
		Phototherapy Radiometer Manufacturer:Fluke Model:DALE 40 SN:2583058

Hany El Desouki

Rasna

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt

Tel.: (202) 25275220/5/6/7

Fax: (202) 25275224

F10P15G

5.2/Oct 2015

Page 1 of 1

كورنيش المادي - برج رياض المادي - القاهرة - مصر

التيلفون : ٢٥٢٧٥٢٢٠ / ٥ / ٦ / ٧ (٢٠٢)

فاكس : ٢٥٢٧٥٢٢٤ (٢٠٢)