

Schedule of Accreditation
for Testing Laboratory According to ISO/IEC 17025
Issued To
Central Laboratories
Alexandria Sanitary Drainage Company
Infront of 21 Moh. Shafik Ghorball St.
Shatby –Alex

Schedule No.: 217021 B

Issue No. (1): May10, 2018

1st Accreditation date: May10, 2018

Revision No. ():

Valid to: October 31, 2020

Material / Products Tested	Type of Tests / Properties Measured / Range of Measurements	Standard Specifications / Techniques Used	
Water and Wastewater	Determination of pH	SMWW 4500-H ⁺ B:2012	pH Meter Model HANNA HI-253 S.N 288808
	Determination of total suspended solid (TSS)	SMWW 2540D:2012	Oven Model:Jeio Tech S.N: N080179 Balance Model : Radwag S.N : 365702/12
	Determination of COD LOQ 23.63 mg/l	SMWW 5220 D:2012	Spectrophotometer Model DR3900-Lange/HACH S.N 1527841

Kornish El-Maadi, Riad El-Maadi Tower 1 - Cairo - Egypt

Tel.: (202) 25275220/5/6/7

Fax: (202) 25275224

F10P15G

5.3/Mar 2018

Page 1 of 1

Hany El Desoky

زينش المعادى - برج رياض المعادى ١ - القاهرة - مصر

نون : ٢٥٢٧٥٢٢٠/٥/٦/٧ (٢٠٢)

س : ٢٥٢٧٥٢٢٤ (٢٠٢)